

PRÜFBESCHEINIGUNG

Die IBExU[®] Institut für Sicherheitstechnik GmbH hat nachfolgende Prüfung als fachkompetentes Unternehmen auf dem Gebiet des Explosionsschutzes durchgeführt.

Hersteller

FLEXIM Flexible Industriemess-
technik GmbH

**Anschrift**

Boxberger Straße 4
12681 Berlin | Deutschland

Produkt

Refraktometer

**Produktprüfung erfolgte
experimentell nach**

EN 60529:1991 + A1:2000 + A2:2013

**Prüfbericht
Prüfergebnis**

IB-21-8-0068 vom 2021-09-09

IP66 1); IP67	PIOX R400, PIOX R500
--------------------------	----------------------

- 1) Die IPX6-Prüfungen der Refraktometer wurden nicht von IBExU durchgeführt, sondern die Ergebnisse wurden aus dem Prüfbericht der Firmen AUCOTEAM GmbH übernommen.

Datum der Ausstellung

2021-09-09

Nummer der Bescheinigung

2103

Ausgestellt von

IBExU[®] Institut für Sicherheitstechnik GmbH
Fuchsmühlenweg 7
09599 Freiberg | Germany

Hinweis 1: Die Bescheinigung wurde maschinell erstellt und ist ohne Unterschrift gültig. Bescheinigungen dürfen nur vollständig und unverändert vervielfältigt werden.

Hinweis 2: Diese Prüfbescheinigung ist nicht zu verwechseln mit einer Bescheinigung im rechtlich geregelten Bereich und darf von dessen Inhaber nicht irreführend oder in unzulässiger Weise verwendet werden.

TEST CERTIFICATE

IBExU[®] Institut für Sicherheitstechnik GmbH has carried out the following examination as a competent company in the field of explosion protection.

Manufacturer

FLEXIM Flexible Industriemess-
technik GmbH

**Address**

Boxberger Straße 4
12681 Berlin | Deutschland

Product

Refractometer

Product testing was carried out experimentally according to

EN 60529:1991 + A1:2000 + A2:2013

Test report

IB-21-8-0068 from 2021-09-09

Test result

IP66 1); IP67	PIOX R400, PIOX R500
--------------------------	----------------------

- 1) The IPX6 tests of the refractometers were not carried out by IBExU, but the results were taken from the test report of AUCOTEAM GmbH.

Date of issue

2021-09-09

Number of the certificate

2103

Issued by

IBExU[®] Institut für Sicherheitstechnik GmbH
Fuchsmühlenweg 7
09599 Freiberg | Germany

Note 1: This test certificate was created automatically and is valid without signature. Certificates may only be reproduced in full and unchanged.

Note 2: This test certificate is not to be used in confusion with a certificate in the legally regulated area and must not be used by the holder in a misleading or impermissible manner.